



UNITED STATES PATENT AND TRADEMARK OFFICE

UNITED STATES DEPARTMENT OF COMMERCE
United States Patent and Trademark Office
Address: COMMISSIONER FOR PATENTS
P.O. Box 1450
Alexandria, Virginia 22313-1450
www.uspto.gov

APPLICATION NO.	FILING DATE	FIRST NAMED INVENTOR	ATTORNEY DOCKET NO.	CONFIRMATION NO.
10/695,632	10/27/2003	Brian G. Connor	S1192.70038US01	6413

7590 01/22/2007
Lawrence M. Green
Wolf, Greenfield & Sacks, P.C.
600 Atlantic Avenue
Boston, MA 02210

EXAMINER

DOWE, KATHERINE MARIE

ART UNIT	PAPER NUMBER
----------	--------------

3734

SHORTENED STATUTORY PERIOD OF RESPONSE	MAIL DATE	DELIVERY MODE
31 DAYS	01/22/2007	PAPER

Please find below and/or attached an Office communication concerning this application or proceeding.

If NO period for reply is specified above, the maximum statutory period will apply and will expire 6 MONTHS from the mailing date of this communication.

Office Action Summary	Application No. 10/695,632	Applicant(s) CONNOR ET AL.	
	Examiner Katherine M. Dowe	Art Unit 3734	

-- The MAILING DATE of this communication appears on the cover sheet with the correspondence address --

Period for Reply

A SHORTENED STATUTORY PERIOD FOR REPLY IS SET TO EXPIRE 1 MONTH(S) OR THIRTY (30) DAYS, WHICHEVER IS LONGER, FROM THE MAILING DATE OF THIS COMMUNICATION.

- Extensions of time may be available under the provisions of 37 CFR 1.136(a). In no event, however, may a reply be timely filed after SIX (6) MONTHS from the mailing date of this communication.
- If NO period for reply is specified above, the maximum statutory period will apply and will expire SIX (6) MONTHS from the mailing date of this communication.
- Failure to reply within the set or extended period for reply will, by statute, cause the application to become ABANDONED (35 U.S.C. § 133). Any reply received by the Office later than three months after the mailing date of this communication, even if timely filed, may reduce any earned patent term adjustment. See 37 CFR 1.704(b).

Status

- 1) ☒ Responsive to communication(s) filed on 27 October 2003.
- 2a) ☐ This action is **FINAL**. 2b) ☒ This action is non-final.
- 3) ☐ Since this application is in condition for allowance except for formal matters, prosecution as to the merits is closed in accordance with the practice under *Ex parte Quayle*, 1935 C.D. 11, 453 O.G. 213.

Disposition of Claims

- 4) ☒ Claim(s) 1-74 is/are pending in the application.
- 4a) Of the above claim(s) _____ is/are withdrawn from consideration.
- 5) ☐ Claim(s) _____ is/are allowed.
- 6) ☐ Claim(s) _____ is/are rejected.
- 7) ☐ Claim(s) _____ is/are objected to.
- 8) ☒ Claim(s) 1-74 are subject to restriction and/or election requirement.

Application Papers

- 9) ☐ The specification is objected to by the Examiner.
- 10) ☐ The drawing(s) filed on _____ is/are: a) ☐ accepted or b) ☐ objected to by the Examiner.
Applicant may not request that any objection to the drawing(s) be held in abeyance. See 37 CFR 1.85(a).
Replacement drawing sheet(s) including the correction is required if the drawing(s) is objected to. See 37 CFR 1.121(d).
- 11) ☐ The oath or declaration is objected to by the Examiner. Note the attached Office Action or form PTO-152.

Priority under 35 U.S.C. § 119

- 12) ☐ Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f).
- a) ☐ All b) ☐ Some * c) ☐ None of:
1. ☐ Certified copies of the priority documents have been received.
2. ☐ Certified copies of the priority documents have been received in Application No. _____.
3. ☐ Copies of the certified copies of the priority documents have been received in this National Stage application from the International Bureau (PCT Rule 17.2(a)).

* See the attached detailed Office action for a list of the certified copies not received.

Attachment(s)

- | | |
|--|---|
| 1) <input type="checkbox"/> Notice of References Cited (PTO-892) | 4) <input type="checkbox"/> Interview Summary (PTO-413)
Paper No(s)/Mail Date. _____ |
| 2) <input type="checkbox"/> Notice of Draftsperson's Patent Drawing Review (PTO-948) | 5) <input type="checkbox"/> Notice of Informal Patent Application |
| 3) <input type="checkbox"/> Information Disclosure Statement(s) (PTO/SB/08)
Paper No(s)/Mail Date _____ | 6) <input type="checkbox"/> Other: _____ |

DETAILED ACTION

1. Restriction to one of the following inventions is required under 35 U.S.C. 121:
 - I. Claims 1-66, 68-69, and 71-74, drawn to a surgical instrument.
 - II. Claims 67 and 70, drawn to a method for performing a surgical procedure on a patient.

The inventions are distinct, each from the other because of the following reasons:

2. Inventions I and II are related as product and process of use. The inventions can be shown to be distinct if either or both of the following can be shown: (1) the process for using the product as claimed can be practiced with another materially different product or (2) the product as claimed can be used in a materially different process of using that product. See MPEP § 806.05(h). In the instant case the method may be used with a curette containing a sharp edge or a curette containing jaws or serrations.

3. This application contains claims directed to the following patentably distinct species:

1) Figs 2-11: Jet-assisted curette:

- 1a) Figs 2-5: Nozzle positioned within cup, cup edge sharpened to form non-liquid jet cutting component
- 1b) Figs 6-7: Sleeve component (606) contains evacuation lumen and high-pressure lumen such that they are separable from curette cup, cup edge sharpened to form non-liquid jet cutting component
- 1c) Figs 8-11: sleeve component contains evacuation lumen and high-pressure lumen such that they are separable from curette cup, cup edge

Art Unit: 3734

comprises claws or serrations, evacuation lumen includes necked-down region

→Embodiments of adjustment of longitudinal position of nozzle and evacuation lumen opening with respect to distal end of cup:

1c-1) Fig 10: sleeve component longitudinally movable while liquid jet length fixed

1c-2) Fig 11: sleeve component position fixed while evacuation lumen and/or high-pressure lumen is moveable with respect to sleeve, allowing liquid jet length to vary

2) Figs 13-17: Jet-assisted rongeur-type instrument: liquid jet fragments, disaggregates, and/or cuts tissue by directly impacting tissue with jet

2a) Figs 13-14: provide nozzle at distal end

2b) Figs 15-17: nozzle insert comprising a nozzle ring with a flat distal-facing surface and a hemispherical proximal-facing surface

3) Figs 18-20: Jet-assisted rongeur-type instrument: liquid jet fragments, disaggregates, and/or cuts tissue indirectly by driving rotor coupled to a rotatable tissue contacting component

3a) Fig 19: Debris evacuated via evacuation lumen (1835) centrally located along the length of hollow burr shaft (1824)

3b) Fig 20: Debris evacuated through the annular space (2004) surrounding burr shaft and within burr shaft; rotating burr shaft (2002) is solid and does not include evacuation lumen

Art Unit: 3734

4. The species are independent or distinct because the species claimed are not obvious variants and are not capable of use together.

Applicant is required under 35 U.S.C. 121 to elect a single disclosed species for prosecution on the merits to which the claims shall be restricted if no generic claim is finally held to be allowable. Currently, no claims are generic.

Applicant is advised that a reply to this requirement must include an identification of the species that is elected consonant with this requirement, and a listing of all claims readable thereon, including any claims subsequently added. An argument that a claim is allowable or that all claims are generic is considered nonresponsive unless accompanied by an election. Applicant must elect invention I or invention II. Within the elected invention, applicant must elect one species 1, 2, or 3 and within the elected species, applicant must further elect one species a, b, or c. If species 1c is elected, applicant must additionally elect either 1c-1 or 1c-2.

Upon the allowance of a generic claim, applicant will be entitled to consideration of claims to additional species which depend from or otherwise require all the limitations of an allowable generic claim as provided by 37 CFR 1.141. If claims are added after the election, applicant must indicate which are readable upon the elected species.

MPEP § 809.02(a).

5. A telephone call was made to Michael Pomianek on January 11, 2007 to request an oral election to the above restriction requirement, but did not result in an election being made.

6. Applicant is advised that the reply to this requirement to be complete must include (i) an election of a species or invention to be examined even though the requirement be traversed (37 CFR 1.143) and (ii) identification of the claims encompassing the elected invention.

The election of an invention or species may be made with or without traverse. To reserve a right to petition, the election must be made with traverse. If the reply does not distinctly and specifically point out supposed errors in the restriction requirement, the election shall be treated as an election without traverse.

Should applicant traverse on the ground that the inventions or species are not patentably distinct, applicant should submit evidence or identify such evidence now of record showing the inventions or species to be obvious variants or clearly admit on the record that this is the case. In either instance, if the examiner finds one of the inventions unpatentable over the prior art, the evidence or admission may be used in a rejection under 35 U.S.C.103(a) of the other invention.

7. Applicant is reminded that upon the cancellation of claims to a non-elected invention, the inventorship must be amended in compliance with 37 CFR 1.48(b) if one or more of the currently named inventors is no longer an inventor of at least one claim remaining in the application. Any amendment of inventorship must be accompanied by a request under 37 CFR 1.48(b) and by the fee required under 37 CFR 1.17(i).

Conclusion

8. Any inquiry concerning this communication or earlier communications from the examiner should be directed to Katherine M. Dowe whose telephone number is (571)272-3201. The examiner can normally be reached on M-F 8:30am - 5pm.

If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, Michael J. Hayes can be reached on (571)272-4959. The fax phone number for the organization where this application or proceeding is assigned is 571-273-8300.

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see <http://pair-direct.uspto.gov>. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free). If you would like assistance from a USPTO Customer Service Representative or access to the automated information system, call 800-786-9199 (IN USA OR CANADA) or 571-272-1000.

Katherine Dowe
January 11, 2007



MICHAEL J. HAYES
SUPERVISORY PATENT EXAMINER